

**RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY, JAIS, AMETHI**

**CORRIGENDUM-1**

**Date: 15.11.2019**

**Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/47 in rgipt.euniwizarde.com portal & Institute Website for the procurement of CHNS.**

**Corrections/Addendum to CHNS analyser Technical Specifications**

<b><u>Before Pre-bid meeting</u></b>	<b><u>After Pre-bid meeting</u></b>
Compliance: Should conform to ASTM D 5373-93, ASTM D5291-91, ASTM D 1552-16, ASTM D 4239	Compliance: Should conform to ASTM D 5373-93, ASTM D5291-91, ASTM D 1552-16
Analysis range: 0.01-100% for all Five elements <ul style="list-style-type: none"><li>➤ From 0.01 mg-100mg or more for C absolute or better</li><li>➤ From 0.01 mg-10mg or more for Hydrogen absolute or better</li><li>➤ From 0.01 mg-100mg or more for Nitrogen absolute or better</li><li>➤ From 0.01 mg-10 mg or more for Sulphur absolute or better</li></ul>	Analysis range: 0.01-100% for all Four elements <ul style="list-style-type: none"><li>➤ From 0.01 mg- 3 mg or more for C absolute or better</li><li>➤ From 0.01 mg- 1 mg or more for Hydrogen absolute or better</li><li>➤ From 0.01 mg- 2 mg or more for Nitrogen absolute or better</li><li>➤ From 0.01 mg- 2 mg or more for Sulphur absolute or better</li></ul>
Balance: Suitable electronic balance of 0.01 mg accuracy	Balance: Suitable electronic balance of 0.001 mg accuracy

**N.B.: A marking system will be used to evaluate the bids. The bid that receives more than the threshold mark fixed by the purchase committee will be considered further for finance bid.**

**The last date to receive the bids is extended till 28<sup>th</sup> Nov 2019 3pm & the Opening of Techno- Commercial bid will be at 4pm.**

  
Dr. U. Ojha

Chairman, Institute Equipment Purchase Committee

  
M.S. Singh