

**RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY  
JAIS, AMETHI**

**Corrigendum - 1**

**Dated:** November 29, 2021

Please refer to the E-tender No. RGIPT/JAIS/R&D/P-2109/SK/2021-22/01 in [rgipt.euniwizarde.com](http://rgipt.euniwizarde.com) portal & Institute Website for the procurement of "01 no. Flow Loop Setup".

Corrections/Addendum to "Flow Loop Setup" Technical Specification

S. No.	Tender Doc. Page No.	Item/ component	Earlier Technical Specifications	Modified Technical Specifications
1	4	Pump	<ul style="list-style-type: none"><li>- Suitable for recirculation of viscous heavy/waxy crude oils, with pulse damper or without any pulsation</li><li>- Digital display</li><li>- Flow rate: 5 to 40 ml/min, or better</li><li>- Ability to withstand minimum 450 psi pressure, or better</li></ul>	<ul style="list-style-type: none"><li>- Suitable for recirculation of viscous heavy/waxy crude oils (<i>having viscosity up to 200 cp or more at 65 °C</i>), with pulse damper or without any pulsation</li><li>- Digital display</li><li>- Flow rate: 10 to 40 ml/min, or better</li><li>- Maximum delivery pressure: atm to 450 psi, or better</li></ul>
2	4	Line heaters	<ul style="list-style-type: none"><li>- To control and indicate the temperature of flow line</li></ul>	<ul style="list-style-type: none"><li>- To control and indicate the temperature of flow line</li><li>- Maximum temperature rating: 100 °C, <i>or better</i></li></ul>
3	5	Electronic control, real time data acquisition and local display	Complete setup must be electronically controlled with power supply and should have the provision of local display of all associated data (all temperature, pressure and flow data).	Complete setup must be electronically controlled with power supply and should have the provision of local display of all associated data (all temperature, pressure and flow data). (Note: Valves can be manually/ electronically controlled)

The last date to receive the bid is extended till **December 13<sup>th</sup>, 2021 (Monday) 3:00 PM** & the Opening of the Techno-commercial bid will be at 4 PM.



Sd/-  
Chairman Purchase Committee