

RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY, JAIS, AMETHI

CORRIGENDUM-1

Date: 15.11.2019

Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/44 in rgipt.euniwizarde.com portal & Institute Website for the procurement of DIFFERENTIAL SCANNING CALORIMETRY (DSC).

Corrections/Addendum to DSC Technical Specifications

<u>Before Pre-bid meeting</u>	<u>After Pre-bid meeting</u>
Cryothermostat: less than equal to -70 °C to 200 °C or better	Cryothermostat: less than equal to -90 to 500 °C or better
Baseline reproducibility : < 1.0 mW or better	Baseline reproducibility : < 50 μW or better
DSC sensitivity (absent)	DSC sensitivity 0.2 μW

N.B.: A marking system will be used to evaluate the bids. The bid that receives more than the threshold mark fixed by the purchase committee will be considered further for finance bid.

The last date to receive the bids is extended till 28th Nov 2019 3pm & the Opening of Techno- Commercial bid will be at 4pm.

Dr. U. Ojha

Chairman, Institute Equipment Purchase Committee

M.S. 