

RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY, JAIS, AMETHI

CORRIGENDUM-1

Date: 27.06.2019

Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/15 in rgipt.euniwizarde.com portal & Institute Website for the procurement of 01 No. Homogenizer.

Name of Equipment	01 No. Homogenizer
Submission Date	28.06.2019 (Friday) before 3pm
Date & Time of Opening of Techno- Commercial bid	28.06.2019 (Friday) at 4pm
Revised Submission Date	15.07.2019 (Monday) before 3pm
Revised Date & Time of Opening of Techno- Commercial bid	15.07.2019 (Monday) at 4pm

A. Technical Specifications for Homogenizer

- 1) Motor output: less than or equal to 100 W
- 2) **Maximum volume: Up to 4 Litre or better**
- 3) Minimum volume: 2 mL
- 4) **Speed range: upto 30,000 rpm**
- 5) Extension arm diameter: smaller than 10 mm.
- 6) Extension arm length: greater than 80 mm
- 7) **Weight: up to 2.0 kg.**
- 8) Voltage: 230 V, 50 Hz
- 9) Maximum noise level: 70 dB(A)
- 10) **Display for rpm**

B. On-site Installation and Training free

C. Two years on-site warranty with spares.

Note: Highlighted portions are the changes in conditions. The other terms & conditions remain unchanged.

This issues with the approval of Competent Authority.

Sd/-

Assistant Registrar (Accounts)