



RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY

[An Institute of National Importance established under an Act of Parliament]

Mubarakpur, Mukhetia, Bahadurpur Post: Harbanshganj, Jais, Amethi Pin Code- 229 304

Website: www.rgipt.ac.in

E-Procurement of Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System

| Type of Tender | E-TENDER (OPEN) |
|--|--|
| E-Tender No. | RGIPT/JAIS/CRF/E-OPN/LAB/2020-21/01 |
| Item Description | Fully PC controlled simultaneous reading Polychromator based Inductively Coupled Plasma Optical Emission Spectrophotometer System (ICP-OES system) with accessories |
| Tender Type | Two-Part Bid System: Tender should be submitted on the schedule to tender. (i) Techno-Commercial Bid and (ii) Financial Bid The Financial bid will be opened only of those firms, who will be found technically qualified after evaluation of their technical bids. |
| Earnest Money Deposit (EMD) | Rs.1,80,000/- (Rupees One Lakh Eighty Thousand only) – Through online mode only. |
| Mode of Submission | Online Mode |
| E-Tender Date | 18.11.2020 (Wednesday) |
| Pre-Bid Meeting | 25.11.2020 (Wednesday) 2.30pm |
| Last date of submission of E-Tender | 09.12.2020 (Wednesday) before 3.00 pm |
| Date & Time of Opening of Techno- Commercial bid | 09.12.2020 (Wednesday) at 4.00 pm |
| Date & Time of Opening of Financial Bid | Will be announced later |
| Place of Opening of E-Tender | Conference Room, 2nd Floor, Administrative Building, Rajiv Gandhi Institute of Petroleum Technology, Jais, Amethi, UP |
| E-Tender Processing Fee (ITI Limited) | Rs. 4500 /- + (Applicable GST @18%) through e-payment Gateway available on https://rgipt.euniwizarde.com |
| Bid Address to | The Director Rajiv Gandhi Institute of Petroleum Technology, Mubarakpur, Mukhetia, Bahadurpur |

| | |
|-------------------------|---|
| | POST: Harbanshganj, Jais, Amethi – 229304 |
| Technical Clarification | Name: Dr. Umapasana Ojha Contact No. 9451959597(Mob.) +91 535 270 4510(LL) Email: uojha@rgipt.ac.in Name: Dr. Gunjan Kumar Agrahari Email: gagrahari@rgipt.ac.in Contact Nos. 9775885136(Mob.) +91 535 270 4514(LL) Department: Chemical Engineering & Engineering Sciences |

Tenders can be downloaded from www.rgipt.ac.in and <https://rgipt.euniwizarde.com/>

E Tendering Helpdesk Number: 011-49606060/8448288988/9650970101/8210817180

E Tendering Helpdesk email ID: ewizardhelpdesk@gmail.com,
ewizardhelpdesk89@gmail.com

INSTRUCTIONS FOR BIDDERS

The bidders are required to submit soft copies of their bids electronically on the e-tender Portal, using valid class 3 Digital Signature Certificates. The instructions given below are meant to assist the bidders in registering on the e-tender Portal, prepare their bids in accordance with the requirements and submitting their bids online on the e-tender Portal.

More information useful for submitting online bids on the e-tender Portal may be obtained at: <https://rgipt.euniwizarde.com>.

REGISTRATION

1. Bidders are required enroll on the e-Procurement Portal (URL: <https://rgipt.euniwizarde.com>) with clicking on the link "Online bidder Registration" on the e-tender Portal by paying the Registration fee of Rs.2360/- Per vendor/per year.
2. As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.
3. Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication with the bidder.
4. Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate (**Only Class III Certificates with signing + encryption key usage**) issued by any Certifying Authority recognized by CCA India (e.g. Sify / TCS / nCode / eMudhra etc.), with their profile.
5. Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSC's to others which may lead to misuse.
6. Bidder then logs in to the site through the secured log-in by entering their user ID /password and the password of the DSC / e-Token.
7. **The scanned copies of all original documents should be uploaded on portal <https://rgipt.euniwizarde.com>**

SEARCHING FOR TENDER DOCUMENTS

1. There are various search options built in the e-tender Portal, to facilitate bidders to search active tenders by several parameters.
2. Once the bidders have selected the tenders they are interested in, you can pay the form fee and processing fee (NOT REFUNDABLE) by net-banking / Debit / Credit card then you may download the required documents / tender schedules, Bid documents etc. Once you pay both fee tenders will be moved to the respective 'requested' Tab. This would enable the e- tender Portal to intimate the bidders through SMS / e-mail in case there is any corrigendum issued to the tender document.

PREPARATION OF BIDS

1. Bidder should take into account any corrigendum published on the tender document before submitting their bids.

2. Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid.
3. Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF formats. Bid Original documents may be scanned with 100 dpi with Colored option which helps in reducing size of the scanned document.
4. To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use “My Documents” available to them to upload such documents.
5. These documents may be directly submitted from the “My Documents” area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process.

SUBMISSION OF BIDS

1. Bidder should log into the website well in advance for the submission of the bid so that it gets uploaded well in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.
2. The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document as a token of acceptance of the terms and conditions laid down by RGIPT, Jais.
3. Bidder has to select the payment option as “e-payment” to pay the tender fee / EMD as applicable and enter details of the instrument.
4. Bidder should prepare the EMD as per the instructions specified in the tender document.
5. Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. If the price bid has been given as a standard BOQ format with the tender document, then the same is to be downloaded and to be filled by all the bidders. Bidders are required to download the BOQ file, open it and complete the white Colored (unprotected) cells with their respective financial quotes and other details (such as name of the bidder). No other cells should be changed. Once the details have been completed, the bidder should save it and submit it online, without changing the filename. If the BOQ file is found to be modified by the bidder, the bid will be rejected.
6. The server time (which is displayed on the bidders’ dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.
7. All the documents being submitted by the bidders would be encrypted using PKI encryption techniques to ensure the secrecy of the data. The data entered cannot be viewed by unauthorized persons until the time of bid opening. Data storage encryption of sensitive fields is done. Any bid document that is uploaded to the server is subjected to symmetric encryption using a system generated symmetric key. Further this key is subjected to asymmetric encryption using buyers/bid opener public keys. Overall, the uploaded tender documents become readable only after the tender opening by the authorized bid openers.

8. The uploaded tender documents become readable only after the tender opening by the authorized bid openers.
9. Upon the successful and timely submission of bid click “Complete” (i.e. after Clicking “Submit” in the portal), the portal will give a successful Tender submission acknowledgement & a bid summary will be displayed with the unique id and date & time of submission of the bid with all other relevant details.
10. The tender summary has to be printed and kept as an acknowledgement of the submission of the tender. This acknowledgement may be used as an entry pass for any bid opening meetings.

ASSISTANCE TO BIDDERS

1. Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.
2. Any queries relating to the process of online bid submission or queries relating to e-tender Portal in general may be directed to the 24x7 Helpdesk Support. **The contact number for the helpdesk is “011-49606060/ 8448288988/ 9650970101” and E Tendering Helpdesk email ID: ewizardhelpdesk@gmail.com**

E-Procurement of Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System

Online Tenders from reputed Original Equipment Manufacturers/ Authorized Dealers for cumulative procurement of **Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System** are invited by the Rajiv Gandhi Institute of Petroleum Technology, Amethi- 229 304 in **Two-Part Bid System** as per specifications mentioned below:

ITEM PARTICULARS

TECHNICAL SPECIFICATIONS

Tender No: - RGIPT/JAIS/CRF/E-OPN/LAB/2020-21/01

Date: 18.11.2020

Fully PC controlled simultaneous reading Polychromator based Inductively Coupled Plasma Optical Emission Spectrophotometer System (ICP-OES system) using high performance solid-state detector technology.

| S. No. | Item | Detailed Specification |
|---------------|---------------------------|---|
| 1. | Spectrometer System | <ul style="list-style-type: none"> • The system must be a true simultaneous system with high performance solid state charge coupled device detectors for high speed, continuous wavelength coverage or segmented charge coupled device (SCD) detectors or charge injection device detector systems equipped with high performance solid state camera system. • Should have capability to perform analysis in DUAL VIEW (both Axial and Radial modes) in a single method. • Must be able to read Axial, Radial or Mix view under single analysis program without any manual intervention • Must be equipped with Polychromator for the best accuracy and precision with fast startup time for low gas consumption. • Startup-time should be less than 30 minutes from instrument standby mode. Minimum Integration Time/read time should be 1 second or better • Must be equipped with Echelle-based simultaneous optics and solid state detector started at -15°C or better • Should have vertical orientation of plasma torch • Full coverage of wavelength in axial or radial view without any loss of sensitivity or light. It should allow choice of wavelength scanning in between the entire wavelength range to cover all elements. The wavelength spectrum should be from 170-780 nm or better to cover all elements possible by ICP-OES. • Should utilize no Argon gas during standby or switch off mode. |
| 2. | Radio Frequency Generator | <ul style="list-style-type: none"> • Should be equipped with solid state free running RF generator running at frequency of 27.12 MHz or 40 MHz with suitable power wattage adjustable from 1000 to 1400 watts or better in 10 watt increment or better. • The power efficiency of RF generator should be greater than 75% with $\leq 1\%$ variation in output power stability. |

| | | |
|----|-------------------------------|--|
| 3. | Plasma View | <ul style="list-style-type: none"> • The system should include complete dual viewing optics under computer and software control. • Any wavelength needed can be used in radial, axial, mixed viewing modes or simultaneous dual view in a single run. • Adjustable radial viewing height of 6 to 18 mm or better |
| 4. | Gas Flow Controls | <ul style="list-style-type: none"> • Should be equipped with computer controlled built-in (3+1) channel electronic mass flow controllers for plasma, auxiliary, nebulizer gas; for precise control of variable gas flow rate (No manual mass flow controller will be accepted). The flow controllers must be covered in warranty, if not, then 2 numbers of extra flow controllers should be provided as spare along with the system. • Plasma gas Flow: 8-20 L/min in 1 L/min increment or suitable range • Auxiliary gas flow: 0-2 L/min in 0.1 L/min increment or suitable range & increment. • Nebulizer gas flow: 0 - 1.5 L/min in 0.01 L/min increment or suitable range & increment. • Additional /makeup gas flow 0.0-0.20 or suitable flow adjustment |
| 5. | Startup and Gas consumption | <ul style="list-style-type: none"> • System should have the least total Argon gas consumption along with least time to perform analysis. (Please mention complete Argon consumption in L/min including Plasma, Auxiliary, Nebulizer, purging gas flows) along with least time to perform analysis & including time for changing any hardware for hydride and non-hydride simultaneous measurement. • Any other gas required should be clearly mentioned and cylinders should be provided. System should have the least warm up time and minimum argon consumption. (Please mention time in minutes from switching off from main switch to the first sample aspiration along with argon gas flow is required during warmup time along with argon consumption rate with respect to analysis timing and any hardware changeover. Three replicates after the first aspiration should be demonstrated with satisfactory repeatability to ensure stability of plasma and suitable warm up time.) • System should have the maintenance free provision to remove the tail plume of plasma. If the system requires any consumable parts. Kindly include the same in warranty period (min 3). OEM noise free compressor should be offered (if applicable) |
| 6. | Hydride Generator | <ul style="list-style-type: none"> • A Hydride generator kit for hydride forming elements like As, Hg, Se etc. should be included with the ICP-OES system. Separate spray chamber & tubing set should be offered. |
| 7. | Detector and wavelength range | <ul style="list-style-type: none"> • Should be equipped with solid state charge coupled device (CCD) /charge injection device (CID) etc. • The actual resolution (not the pixel resolution) of the |

| | | |
|-----|---|--|
| | | <p>system must be less than 0.007 nm at 200/202 nm or better.</p> <ul style="list-style-type: none"> • The spectrophotometer must cover full spectral range from 170-780 nm or suitable range to cover all elements possible by ICPOES. • The system should have capability to measure in between 2000- 10,000 emission wavelengths. • Should have simultaneous <i>UV</i> and <i>Vis</i> analyses • It should allow intense and trace signals to be measured simultaneously at the optimum signal to noise ratio. • Should be able to determine major, minor, and trace elements in single run, and should be able to detect minimum concentration of one parts per billion (1 ppb) or even better |
| 8. | Sample Introduction system | <ul style="list-style-type: none"> • ICP-OES system should have integrated sample introduction system with minimum four channels computer controlled peristaltic pump with variable speed from 0.2 ml to 7 mL per minute or 0-80 rpm or better, for maximum flexibility. The pump allows for sample, drain, internal standard, ionization buffer, reducing agents etc. • The sample introduction system should include concentric nebulizer, spray chamber, torch, and injector. Should be made up of highly resistant material to take care of various acids like HF, HCl, HNO₃, HClO₄ etc. • One piece quartz torch, double-pass/cyclonic spray chamber, low flow concentric glass nebulizer should be standard supply with capability of dual view radial and axial. • Built-in sprint valve/automated valve system to have fast sample analysis be offered |
| 9. | TDS capability | <ul style="list-style-type: none"> • ICP-OES system should allow analysis of samples with high TDS up to 3%. Separate High TDS kit to be included in the offer. |
| 10. | Background correction and Spectral interference | <ul style="list-style-type: none"> • ICP-OES system should have the facility of online/simultaneous background correction. • The ICP-OES system must be able to read and apply manual or automatic spectral interferences correction in addition to background correction |
| 11. | Instrument Software | <ul style="list-style-type: none"> • The instrument handling original software should be based on <i>Windows-10 Professional</i> installed on the computer with color printer. |
| 12. | Sample Digestion System: Microwave Digester | <ul style="list-style-type: none"> • Dual magnetron based system must be capable of processing 8 or more reaction vessels. • Specifications of Microwave Digester: <ul style="list-style-type: none"> • Volume of reaction vessel: 80 ml or more • Material of reaction vessel: Teflon Vessels • Number of reaction vessel: Eight (08) • Operation temperature: 240⁰C or above • Maximum Vessel Temperature 300⁰C • Operation pressure: 60 bar or more |

| | | |
|-----|--------------------------------------|---|
| | | <ul style="list-style-type: none"> • Power requirement: 1700 W or more • Current requirement: 220-240V, v50 Hz • Pressure sensors/probes to control the pressure in each reaction vessels • Temperature sensors/probes for temperature control • Software supported • Material of Construction: stainless steel |
| 13. | Accessories, spares, and consumables | <ul style="list-style-type: none"> • One set of inert torch and accessories required for running HF digested samples. • Sampling kits for aqueous solutions • Suitable Chiller re-circulator of appropriate capacity along with the system. • Fume hood: Vendor should quote appropriate fume hood complete including exhaust fan for the ICP system. • Vendor must quote filled Argon gas cylinders (05 numbers) with double stage gas pressure regulator (01 no.) and purification panels, nitrogen cylinders (no. 2) with pressure regulator. • OEM noise free Air compressor, if required • Computer & Printer: The system should have appropriate branded workstation PC with minimum core-i7 processor, 8 GB RAM, and Deskjet color printer • Two numbers of reaction crucible made of platinum of volume 15 ml or more • Consumables: Sample Tubes: 10 set, Waste/Drain Tubes: 10 set, Torch: 3, Concentric Nebulizer: 2, Injector: 2, Spray Chamber-1 • Operation and maintenance manual should be provided in both hard and soft copies. • Pre installation requirements- Complete technical details of pre installation requirement should be furnished along with the technical bid. |
| 14. | Standards | <ul style="list-style-type: none"> • 200 mL or more of Multi Element Standards (NIST standard containing 23 elements) of 1000 ppm each element with traceability certificate from competent agency containing transition elements, alkali elements, alkaline earth elements. • Separate 100 ppm standards for Hg, Se, Sn (100 mL) |
| 15. | General | <ul style="list-style-type: none"> • Chiller/re-circulator should not be mounted to the instrument chassis and should be from the original manufacturer of imported origin |
| 16. | System software | <ul style="list-style-type: none"> • The instrument system software shall be based on the Windows-10 operating system. • Software should also have comprehensive wavelength library with indication of preferred line for each element. • It should feature automatic identification of possible spectral interferences when selecting wavelengths for analysis and should have search mode for identification of unknown wavelengths. |
| 17. | Warranty and | <ul style="list-style-type: none"> • Comprehensive warranty for 2 years apart from standard |

| | | |
|--|-------------|--|
| | Maintenance | warranty of first year including Spares and Consumables must be included <ul style="list-style-type: none"> • Warranty should be offered from OEM with Part number etc. |
|--|-------------|--|

Place:

Date:
Bidder

Signature and seal of the Manufacturer/

Technical Specifications Compliance Statement of “**Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System**” should be submitted as per format given in Annexure – I (**On the Company’s Letter Head**)

Terms & Conditions

1. Tender filling up and submission

- a. Tender offer must refer to the letter/Notice inviting the Tender.
- b. The bid should be uploaded online mentioning “Supply & Installation of **“Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System”** to Rajiv Gandhi Institute of Petroleum Technology, Jais, Amethi.
- c. Bids must be submitted online against above mentioned tender.
- d. The prices can be quoted in Indian Rupees or Foreign Currency, both in figures and in words and the same will be considered for price comparison. Where there is a difference between the two, the rates given in words will be taken as authentic. Should there be any discrepancy between unit rate and amount, the unit rate will be considered as the correct one.
- e. The rates quoted should include freight charges upto Jais, Amethi clearly mentioning the percentage/rate of GST and the rates quoted should be valid for at- least 60 days from the date of opening of the quotation. The rates must be quoted both in figures and words and overwriting should be avoided. However all cuttings/corrections must be duly authenticated.
- f. The bid submitted must not contain any ambiguity. Any such ambiguity in the quote submitted shall make it liable for rejection.
- g. The tender document is not transferable.
- h. No amendment to the bid in any form by the bidder will be permitted after opening of the tender. No conditional offers shall be accepted.
- i. Submission of bid will bind the bidder to the acceptance of all conditions specified in the tender document.
- j. Canvassing in any form is for bidden and will be a criterion for disqualification.
- k. **Price bid submission:** In case of imported equipment, price (quote) shall be in two parts – (a) CIF – (Cost, Insurance, Freight) price upto Airport in Foreign currency, and (b) from Airport to destination (RGIPT, Jais, Amethi) in Indian currency. The components covered in this schedule consist of custom duty; custom clearance charges; freight charges, & Installation and training charges etc, which will be reimbursed in Indian currency to the Indian counterpart as per actual (Up to the maximum limit quoted in the bid), on submission of the original supporting bills duly certified and stamped by the Chairman Purchase Committee.
- l. The Purchase Order (PO) to import any equipment shall be placed on the name of ‘foreign principal’ only in foreign currency.

2. Special Instructions:

- i. Non-receipt of quote in foreign currency for imported equipment may lead to the rejection of their bid without assigning any reason thereof.
- ii. RGIPT shall provide custom/excise duty exemption certificate and GST exemption certificate on request by the bidder.
- iii. RGIPT, Jais, Amethi avails custom duty exemption in terms of Government Notification No. 51/96-Customs dated 23 July, 1996, and Central Excise Duty Exemption in terms of Government Notification No. 10/97-Central Excise dated 01 March, 1997, as amended time to time.
- iv. RGIPT, Jais, Amethi avails GST concession under notification No. 47/2017 central tax (rate) dated 14 November 2017, as amended time to time.
- v. The earnest money deposit (EMD)/ bid security shall be submitted in Indian currency only. The bid security should remain valid for a period of 45 days beyond the final bid validity period.
- vi. The company before the opening of letter of credit (LC) or wire transfer shall submit the performance security. The company shall submit the performance security at the

rate of five percent (5%) of contract value in Indian Rupees in the form of demand draft / FDR/PBG, issued in the favor of Rajiv Gandhi Institute of Petroleum Technology” payable at Bank of Baroda, RGIPT, Jais campus branch, Amethi from any of the scheduled bank.

- vii. The performance guarantee should remain valid at least for a period of 60 days of beyond the date of completion of all contractual obligations of supplier.
 - viii. The eighty percent (80%) payment of contract value shall be paid at site by an irrevocable letter of credit, two weeks prior to dispatch of equipment on submission of original documents without any discrepancies and balance twenty percent (20%) shall be released after satisfactory installation and commissioning of equipment.
3. **EMD (Earnest Money Deposit):** The EMD should pay only in online mode on <https://rgipt.euniwizarde.com> of **Rs. 1, 80,000/- (Rupees One Lakh Eighty Thousand only) – Through online mode only**). No interest shall be paid on earnest money deposited.
- a) The EMD of the unsuccessful bidders will be returned after final award of the Job to a successful bidder/ acceptance of the tender and within reasonable time not exceeding six months from the date of opening of Tender. The EMD will carry no interest.
 - b) The EMD will be forfeited if the bidder withdraws his offer before finalization of the tender.
 - c) The EMD of successful bidder will be returned after the submission of Performance Security.
 - d) The firms registered with DGS&D, MSME & NSIC, if any, are exempted from payment of EMD provided such registration includes the item they are offering which are manufactured by them and not for selling products manufactured by other companies. **(Latest Copy must be submitted online for claiming exemption)**.
4. **Performance Security:** The successful bidder will provide “Performance Security” for an amount of 5% of the contract value in the form of valid Account payee Demand Draft/ FDR/PBG from a scheduled bank favoring “Rajiv Gandhi Institute of Petroleum Technology” payable at RGIPT, Jais, Amethi Campus Branch. Performance security should remain valid at least for a period of 180 days beyond the date of completion of all contractual obligations of supplier **(Total Period of FDR/Demand Draft/PBG – Warranty Period plus 180 days)**. Thereafter the security will be returned, provided there is no defect in the equipment supplied.
5. **Evaluation Criteria:** The order will be awarded to the vendor, who fulfills all the technical specifications (Marks based evaluation) & conditions mentioned in the tender document for Technical Bid and whose rate is L1 in grand total in Financial Bid. Evaluation of Lowest bid will be calculated on the total price of all items tendered for basic equipment along with accessories, packing & forwarding, Installation & Commissioning, GST and other charges etc. as mentioned in the tender (Bidder must submit the technical bid & financial bid separately). The Financial bid will be opened only of those firms, who will be found technically qualified after evaluation of their technical bids.
6. **Bid Validity:** The bid must be valid for 180 days from the date of opening of Technical Bid.

7. **Payment terms:**

- a) **In Indian Currency Payment terms** - 80% payment will be released against successful supply of equipment at site i.e. Rajiv Gandhi Institute of Petroleum Technology, Jais, Amethi and balance 20% after installation & commissioning of the equipment and subject to submission of original Bill/ Invoice as certified by the Indenter.
- b) **In Foreign currency payment** -80% payment by an irrevocable letter of credit favouring Principal Company and for compliance of payment term, two weeks prior to dispatch of the equipment on submission of letter from Principal Company confirming that the equipment is ready for dispatch and on receipt of documents without any discrepancies and balance 20% will be paid after satisfactory installation and commissioning as certified by the indenter.

Reimbursement of Expenses in Indian Rupees: The Institute will reimburse the expenses in Indian Rupees related to custom duty, custom clearance charges, transportation charges upto RGIPT, Jais, Amethi to the Indian Counterpart after the installation and commissioning of equipment at site (Upto maximum limit defined in the bid) i.e Rajiv Gandhi Institute of Petroleum Technology, Jais, Amethi – 229304 and subject to submission of original bill certified by the Indenter.

8. **Clearance and delivery:** On arrival of shipment/consignment, the supplier is required to arrange custom clearance and transportation of the consignment up to RGIPT, Jais, Amethi premises. The Tenderer/ bidder will do all types of clearance work and formalities to deliver/provide the equipment at the site of installation. The expenses incurred will be reimbursed on actual basis not exceeding the quoted amount. Institute will provide all types of documentary support including Customs Duty exemption certificate. The custom duty, custom clearance charges, transportation charges etc. will be reimbursed in Indian Rupees on subject of submission of documentary proof. This may be included in your quotation as a separate item. Principals will themselves have to procure any requisite permission from the Govt. of country of origin of equipment.

9. **Part Shipment not allowed:** Part Shipment will NOT be allowed.

10. **Insurance:** For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "Warehouse to Institute" (final destination) on "All Risks" basis including War Risks and Strikes. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. If orders placed on CIF/CIP basis, the insurance should be up to RGIPT, Jais, Amethi.

11. **Bank Charges:** All the bank charges within India will be borne by the Institute and outside India will be borne by the Supplier.

12. **Site Preparation:** The supplier shall inform to the Institute about the site preparation, if any, needed for the installation of equipment, immediately after the receipt of the purchase order. The supplier must provide complete details regarding space and all the other infrastructural requirements needed for the equipment, which the Institute should arrange before the arrival of the equipment to ensure its timely installation and smooth operation thereafter. The supplier shall visit the Institute and see the site where the equipment is to be installed and may offer his advice and render assistance to the Institute in the preparation of the site and other pre-installation requirements.

13. **Resolution of Disputes:** The dispute resolution mechanism to be applied pursuant shall be as follows:
- i. In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to any matter arising out of or connected with this agreement, such disputes or difference shall be settled in accordance with the Indian Arbitration & Conciliation Act, 1996, the rules there under and any statutory modifications or re-enactments thereof shall apply to the arbitration proceedings. The dispute shall be referred to the Director, Rajiv Gandhi Institute of Petroleum Technology, Jais, Amethi and if he is unable or unwilling to act, to the sole arbitration of some other person appointed by him willing to act as such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and binding on all parties to this order.
 - ii. In the case of a dispute between the purchaser and a Foreign Supplier, the dispute shall be settled by arbitration in accordance with provision of sub-clause (a) above. But if this is not acceptable to the supplier then the dispute shall be settled in accordance with provisions of UNCITRAL (United Nations Commission on International Trade Law) Arbitration Rules.
14. Live working product Demo of same or similar item is necessary (at the cost of bidder) for all the bidders, if asked for, within the ONE week of opening the technical bid.
15. **Printed Conditions of the Company – Not Acceptable to the Institute:** RGIPT does not accept the printed conditions of any Tenderer. It will be ignored without any reference; hence tenderers should withdraw such printed conditions if they have any. RGIPT shall accept only unconditional tender.
16. **Technical Deviations – Sole discretion of RGIPT for consideration:** If technical deviations furnished by the Tenderer are not agreeable to RGIPT, the offers may be ignored. However it will be solely at discretion of RGIPT to consider the technical deviations OR not for considering the Tenderer. No correspondences of the Tenderer will be entertained in this matter.
17. **Price Validity:** The price quoted by the bidder shall remain valid till the supply is completed. During the tenure of the supply period, no price increase will be considered, except for change in statutory levies imposed by Centre Govt. / state Govt. or Local authorities.
18. **Warranty:** As specified in the Technical Specification.
19. **Delivery:** The timely delivery of “**Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System**” is the essence of contract. The supply should be completed with the specified delivery. Delivery shall be completed within maximum 16 weeks from the placement of purchase order/from the date of opening of LC.
20. **Bid Acceptance and Rejection:**
- a. RGIPT reserves the right to reject the bid without assigning any reason what so ever.
 - b. RGIPT does not bind itself to accept the lowest or any quotation and reserves the right to reject any or all without assigning any reason.
21. **Late Receipt of Bids:** No bids will accepted or considered after the due date and time.
22. **Liquidity Damage Clause:** In case of any damages and delay in supply of material, liquidated damages at the rate of 0.5% of the order value per day of delay or part thereof,

subject to a maximum of 5% of the order value shall be recovered from the vendor's bill/invoice.

23. **Withdrawal of Tender:** The Institute reserves the right to withdraw the tender at any stage of tendering process with the approval of the Competent Authority.
24. **Replacements of Goods Broken, Damaged or Short:** If whole or part of the goods/ equipment supplied against the order is found defective/ non-functional or without specification the supplier will have to replace the defective goods/equipment to make equipment operational within a reasonable time (not exceeding two months from the date of intimation to the supplier or the Indian Agent), failure to which the supplier will have to refund the full cost of the equipment including freight, Insurance etc.
25. A quotation submitted with false information will not only be rejected but the vendor may also be debarred for participation in future tendering processes.
26. Authorized representative may choose to be present at the time of opening of quotation at their own cost.
27. **Each bidder who will participate in the tender/bid will make sure that each and every item listed in Item Description page in technical specifications of Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System will have to be supplied otherwise their bid will be rejected without any intimation.**
28. **Each bidder should have at least 4 years performance of supplying the Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System in various IIT/NIT/Other National Level Institute in India. PO copy should be attached.**
29. **Only one PRICE BID should be submitted for the above equipment. A bidder who submits more than one PRICE BID, shall be disqualified and considered non-responsive.**

Sd/-
Chairman Purchase Committee

Documents required for eligibility in the tender

| Sl. No. | Description |
|---------|--|
| | The EMD of Rs. 1, 80,000/- (Rupees One Lakh Eighty Thousand only) is to be paid only through online mode on https://rgipt.euniwizarde.com . If the EMD is not paid, then the bid will be automatically rejected. |
| 1 | Technical Specification Compliance Statement -(On the letter head of the Company) - Annexure -1 |
| 2 | Manufacturer's Authorization Form – In defined format OR Self Declaration of Manufacturing Unit – Annexure - 2 |
| 3 | Price Reasonability Certificate – In defined format (On the letter head of the Company)- Annexure - 3 |
| 4 | Declaration Certificate (Acceptance of terms & conditions of the tender) – In defined format (On the letter head of the Company) - Annexure – 4 and Declaration Sheet as per Annexure - 4A . |
| 5 | As on date of submission of the proposal, the Bidder is neither blacklisted by Central Government/ State Government or Instrumentalities thereof nor is any criminal case against the Bidder/ its Partners/ Directors/ Agents pending before any court of law. Non-Blacklisting Certificate on Rs. 100.00 non-Judicial Stamp Paper – In defined format - Annexure - 5 |
| 6 | Copy of Firm Registration |
| 7 | Copy of PAN Card & GST No. |
| 8 | Bidder should have at least 4 years' experience in the selling and 3 years' experience providing service (related Equipment) to reputed Central Government Institutes/Engineering Colleges/ Universities/ Oil Companies/PSUs/Research Institutes/ IITs/ NITs/ IISERs (Self Declaration) – List of Clients & Purchase Orders of “Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System” to be submitted. |
| 9 | The Bidder or their OEM should possess any valid standard certification like ISO 9001:2008 & 14001:2004. |
| 10 | Bidder should enclose the copies of last 3 years Audited Annual Accounts (Balance Sheet and Profit & Loss A/c) duly authenticated by Chartered Accountant |
| 11 | Last three years CA certified turnover certificate (Average Turnover of last three years should be Rs.25lakhs (Rupees Twenty Five lakhs only)) |
| 12 | Last 3 year Income Tax Return (ITR) |
| 13 | In case of exemption of EMD, the scanned copy of the document in support of exemption will have to be uploaded by the bidder during bid submission |
| 14 | MOST IMPORTANT: Document related to prior installation and service – At least 3 Academic and R&D National Organizations, Central and State Government Departments/ Institutions/ PSU in the last 3 years. Annexure - 6 |



RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY

CHECKLIST FOR SUBMISSION OF DOCUMENTS FOR TECHNO-COMMERICAL BID

Part -1

MOST IMPORTANT: PLEASE DON'T ADD OR UPLOAD ANY ANNEXURE OF PRICE BID IN TECHNO-COMMERICAL BID DOCUMENTS, OTHERWISE YOUR BID WILL BE REJECTED.

Checklist for submission of documents for Techno-Commercial Bid

| Sl. No. | Description | Uploaded/ Not Uploaded |
|---------|--|------------------------|
| | The EMD of Rs. 1, 80,000/- (Rupees One Lakh Eighty Thousand only) is to be paid only through online mode on https://rgipt.euniwizarde.com . If the EMD is not paid, then the bid will be automatically rejected. | |
| 1 | Technical Specification Compliance Statement -(On the letter head of the Company) - Annexure - 1 | |
| 2 | Manufacturer's Authorization Form - In defined format OR Self Declaration of Manufacturing Unit - Annexure - 2 | |
| 3 | Price Reasonability Certificate - In defined format (On the letter head of the Company)- Annexure - 3 | |
| 4 | Declaration Certificate (Acceptance of terms & conditions of the tender) - In defined format (On the letter head of the Company) - Annexure - 4 and Declaration Sheet as per Annexure - 4A . | |
| 5 | As on date of submission of the proposal, the Bidder is neither blacklisted by Central Government/ State Government or Instrumentalities thereof nor is any criminal case against the Bidder/ its Partners/ Directors/ Agents pending before any court of law. Non-Blacklisting Certificate on Rs.100.00 non-Judicial Stamp Paper - In defined format - Annexure - 5 | |
| 6 | Copy of Firm Registration | |
| 7 | Copy of PAN Card & GST No. | |
| 8 | Bidder should have at least 4 years' experience in the selling and 3 years' experience providing service (related Equipment) to reputed Central Government Institutes/Engineering Colleges/ Universities/ Oil Companies/PSUs/Research Institutes/ IITs/ NITs/ IISERs (Self Declaration) - List of Clients & Purchase Orders of " Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System " to be submitted. | |
| 9 | The Bidder or their OEM should possess any valid standard certification like ISO 9001:2008 & 14001:2004. | |
| 10 | Bidder should enclose the copies of last 3 years Audited Annual Accounts (Balance Sheet and Profit & Loss A/c) duly authenticated by Chartered Accountant | |
| 11 | Last three years CA certified turnover certificate (Average Turnover of last three years should be Rs.25 lakhs (Rupees Twenty five lakhs only only)) | |
| 12 | Last 3 year Income Tax Return (ITR) | |
| 13 | In case of exemption of EMD, the scanned copy of the document in support of exemption will have to be uploaded by the bidder during bid submission | |
| 14 | MOST IMPORTANT: Document related to prior installation and service - At least 3 Academic and R&D National Organizations, Central and State Government Departments/Institutions/PSU in the last 3 years. Annexure - 6 | |

Place:

Date:

Signature and seal of the Manufacturer/ Bidder

Technical Compliance Statement Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System

(ON THE LETTER HEAD OF THE COMPANY)

Please mention the brand and model no. & name of the equipment:

| S. No. | Item | Detailed Specification | Comply/ Not Comply |
|--------|---------------------------|---|--------------------|
| 1. | Spectrometer System | <ul style="list-style-type: none"> • The system must be a true simultaneous system with high performance solid state charge coupled device detectors for high speed, continuous wavelength coverage or segmented charge coupled device (SCD) detectors or charge injection device detector systems equipped with high performance solid state camera system. • Should have capability to perform analysis in DUAL VIEW (both Axial and Radial modes) in a single method. • Must be able to read Axial, Radial or Mix view under single analysis program without any manual intervention • Must be equipped with Polychromator for the best accuracy and precision with fast startup time for low gas consumption. • Startup-time should be less than 30 minutes from instrument standby mode. Minimum Integration Time/read time should be 1 second or better • Must be equipped with Echelle-based simultaneous optics and solid state detector started at -15°C or better • Should have vertical orientation of plasma torch • Full coverage of wavelength in axial or radial view without any loss of sensitivity or light. It should allow choice of wavelength scanning in between the entire wavelength range to cover all elements. The wavelength spectrum should be from 170-780 nm or better to cover all elements possible by ICP-OES. • Should utilize no Argon gas during standby or switch off mode. | |
| 2. | Radio Frequency Generator | <ul style="list-style-type: none"> • Should be equipped with solid state free running RF generator running at frequency of 27.12 MHz or 40 MHz with suitable power wattage adjustable from 1000 to 1400 watts or better in 10 watt increment or better. • The power efficiency of RF generator should be greater than 75% with $\leq 1\%$ variation in | |

| | | | |
|----|-----------------------------|---|--|
| | | output power stability. | |
| 3. | Plasma View | <ul style="list-style-type: none"> • The system should include complete dual viewing optics under computer and software control. • Any wavelength needed can be used in radial, axial, mixed viewing modes or simultaneous dual view in a single run. • Adjustable radial viewing height of 6 to 18 mm or better | |
| 4. | Gas Flow Controls | <ul style="list-style-type: none"> • Should be equipped with computer controlled built-in (3+1) channel electronic mass flow controllers for plasma, auxiliary, nebulizer gas; for precise control of variable gas flow rate (No manual mass flow controller will be accepted). The flow controllers must be covered in warranty, if not, then 2 numbers of extra flow controllers should be provided as spare along with the system. • Plasma gas Flow: 8-20 L/min in 1 L/min increment or suitable range • Auxiliary gas flow: 0-2 L/min in 0.1 L/min increment or suitable range & increment. • Nebulizer gas flow: 0 - 1.5 L/min in 0.01 L/min increment or suitable rage & increment. • Additional /makeup gas flow 0.0-0.20 or suitable flow adjustment | |
| 5. | Startup and Gas consumption | <ul style="list-style-type: none"> • System should have the least total Argon gas consumption along with least time to perform analysis. (Please mention complete Argon consumption in L/min including Plasma, Auxiliary, Nebulizer, purging gas flows) along with least time to perform analysis & including time for changing any hardware for hydride and non-hydride simultaneous measurement. • Any other gas required should be clearly mentioned and cylinders should be provided. System should have the least warm up time and minimum argon consumption. (Please mention time in minutes from switching off from main switch to the first sample aspiration along with argon gas flow is required during warmup time along with argon consumption rate with respect to analysis timing and any hardware changeover. Three replicates after the first aspiration should be demonstrated with | |

| | | | |
|----|-------------------------------|--|--|
| | | <p>satisfactory repeatability to ensure stability of plasma and suitable warm up time.)</p> <ul style="list-style-type: none"> • System should have the maintenance free provision to remove the tail plume of plasma. If the system requires any consumable parts. Kindly include the same in warranty period (min 3). OEM noise free compressor should be offered (if applicable) | |
| 6. | Hydride Generator | <ul style="list-style-type: none"> • A Hydride generator kit for hydride forming elements like As, Hg, Se etc. should be included with the ICP-OES system. Separate spray chamber & tubing set should be offered. | |
| 7. | Detector and wavelength range | <ul style="list-style-type: none"> • Should be equipped with solid state charge coupled device (CCD) /charge injection device (CID) etc. • The actual resolution (not the pixel resolution) of the system must be less than 0.007 nm at 200/202 nm or better. • The spectrophotometer must cover full spectral range from 170-780 nm or suitable range to cover all elements possible by ICPOES. • The system should have capability to measure in between 2000- 10,000 emission wavelengths. • Should have simultaneous <i>UV</i> and <i>Vis</i> analyses • It should allow intense and trace signals to be measured simultaneously at the optimum signal to noise ratio. • Should be able to determine major, minor, and trace elements in single run, and should be able to detect minimum concentration of one parts per billion (1 ppb) or even better | |
| 8. | Sample Introduction system | <ul style="list-style-type: none"> • ICP-OES system should have integrated sample introduction system with minimum four channels computer controlled peristaltic pump with variable speed from 0.2 ml to 7 mL per minute or 0-80 rpm or better, for maximum flexibility. The pump allows for sample, drain, internal standard, ionization buffer, reducing agents etc. • The sample introduction system should include concentric nebulizer, spray chamber, torch, and injector. Should be made up of highly resistant material to take care of various acids like HF, HCl, HNO₃, HClO₄ etc. • One piece quartz torch, double-pass/cyclonic spray chamber, low flow concentric glass nebulizer should be standard supply with | |

| | | | |
|-----|---|--|--|
| | | <p>capability of dual view radial and axial.</p> <ul style="list-style-type: none"> Built-in sprint valve/automated valve system to have fast sample analysis be offered | |
| 9. | TDS capability | <ul style="list-style-type: none"> ICP-OES system should allow analysis of samples with high TDS up to 3%. Separate High TDS kit to be included in the offer. | |
| 10. | Background correction and Spectral interference | <ul style="list-style-type: none"> ICP-OES system should have the facility of online/simultaneous background correction. The ICP-OES system must be able to read and apply manual or automatic spectral interferences correction in addition to background correction | |
| 11. | Instrument Software | <ul style="list-style-type: none"> The instrument handling original software should be based on <i>Windows-10 Professional</i> installed on the computer with color printer. | |
| 12. | Sample Digestion System: Microwave Digester | <ul style="list-style-type: none"> Dual magnetron based system must be capable of processing 8 or more reaction vessels. Specifications of Microwave Digester: <ul style="list-style-type: none"> Volume of reaction vessel: 80 ml or more Material of reaction vessel: Teflon Vessels Number of reaction vessel: Eight (08) Operation temperature: 240°C or above Maximum Vessel Temperature 300°C Operation pressure: 60 bar or more Power requirement: 1700 W or more Current requirement: 220-240V, v50 Hz Pressure sensors/probes to control the pressure in each reaction vessels Temperature sensors/probes for temperature control Software supported Material of Construction: stainless steel | |

| | | | |
|-----|--------------------------------------|---|--|
| 13. | Accessories, spares, and consumables | <ul style="list-style-type: none"> • One set of inert torch and accessories required for running HF digested samples. • Sampling kits for aqueous solutions • Suitable Chiller re-circulator of appropriate capacity along with the system. • Fume hood: Vendor should quote appropriate fume hood complete including exhaust fan for the ICP system. • Vendor must quote filled Argon gas cylinders (05 numbers) with double stage gas pressure regulator (01 no.) and purification panels, nitrogen cylinders (no. 2) with pressure regulator. • OEM noise free Air compressor, if required • Computer & Printer: The system should have appropriate branded workstation PC with minimum core-i7 processor, 8 GB RAM, and Deskjet color printer • Two numbers of reaction crucible made of platinum of volume 15 ml or more • Consumables: Sample Tubes: 10 set, Waste/Drain Tubes: 10 set, Torch: 3, Concentric Nebulizer: 2, Injector: 2, Spray Chamber-1 • Operation and maintenance manual should be provided in both hard and soft copies. • Pre installation requirements- Complete technical details of pre installation requirement should be furnished along with the technical bid. | |
| 14. | Standards | <ul style="list-style-type: none"> • 200 mL or more of Multi Element Standards (NIST standard containing 23 elements) of 1000 ppm each element with traceability certificate from competent agency containing transition elements, alkali elements, alkaline earth elements. • Separate 100 ppm standards for Hg, Se, Sn (100 mL) | |
| 15. | General | <ul style="list-style-type: none"> • Chiller/re-circulator should not be mounted to the instrument chassis and should be from the original manufacturer of imported origin | |
| 16. | System software | <ul style="list-style-type: none"> • The instrument system software shall be based on the Windows-10 operating system. • Software should also have comprehensive wavelength library with indication of preferred line for each element. • It should feature automatic identification of possible spectral interferences when selecting wavelengths for analysis and should have search mode for identification of unknown wavelengths. | |

| | | | |
|-----|--------------------------|---|--|
| 17. | Warranty and Maintenance | <ul style="list-style-type: none"> • Comprehensive warranty for 2 years apart from one year standard warranty including Spares and Consumables must be included • Warranty should be offered from OEM with Part number etc. | |
|-----|--------------------------|---|--|

Place:

Date:
Manufacturer/Bidder

Signature and seal of the

**MANUFACTURERS' AUTHORIZATION FORM (MAF)
(ON THE LETTER HEAD OF THE COMPANY)**

[NOTE: The Bidder shall require the Manufacturer to fill in this Form in accordance with the instructions indicated. This letter of authorization should be **on the letterhead of the Manufacturer** and should be signed by a person with the proper authority to sign documents that are binding on the Manufacturer. It should be included by the Bidder in its bid]

No. Dated

Date: [insert date (as day, month and year) of Bid Submission]

Tender No.: [insert number from Invitation for Bids]

To, Director
Rajiv Gandhi Institute of Petroleum Technology
Mubarakpur, Mukhtia, Bahadurpur
POST: Harbanshganj, Jais, Amethi - 229304

WHEREAS

We [insert complete name of Manufacturer], who are official manufacturers of [insert type of goods manufactured], having factories at [insert full address of Manufacturers factories], do hereby authorize [insert complete name of Bidder] to submit a bid the purpose of which is to provide the following Goods, manufactured by us [insert name and or brief description of the Goods], and to subsequently negotiate and sign the Contract.

We hereby extend our full guarantee and warranty in accordance with the General Conditions of Contract, with respect to the Goods offered by the above firm.

Signed: [insert signature(s) of authorized representative(s) of the Manufacturer]

Name: [insert complete name(s) of authorized representative(s) of the Manufacturer]

Title/Designation: [insert title]

Duly authorized to sign this Authorization on behalf of: [insert complete name of Bidder]

Dated on ____ day of, _____, _____ [insert date of signing]

Yours faithfully,

(Name of manufacturers)

OR

SELF DECLARATION OF MANUFACTURING UNIT

PRICE REASONABILITY CERTIFICATE
(ON THE LETTER HEAD OF THE COMPANY)

It is certified that the rates quoted against Tender No. datedfor the items vide our bid No. datedare exclusively for supply to ACADEMIC and RESEARCH Institutions and are not more than as charged to other Govt./PSU"s for similar supplies made in recent past. If they have been approved by the Director, RGIPT and if at any stage it has been found that the quoted rates are higher than the rates applicable to supply to Government then in such condition RGIPT, Amethi, will have the right to cancel the approved rates and to take legal action against the tenderer.

Yours faithfully,

(Name of manufacturers)

(Signature with date)

(Name and designation)

Duly authorized to sign tender for and on behalf of

DECLARATION (ACCEPTANCE OF TERMS & CONDITIONS)
(ON THE LETTER HEAD OF THE COMPANY)

1. I, Son/Daughter of Shri
Proprietor/ Partner/CEO/MD/Director/Authorized Signatory of M/s
..... am competent to sign this tender document.
2. I have carefully read and understood all the terms and conditions of the tender and hereby convey my acceptance of the same.
3. The information/ documents furnished along with the above application are true and authentic to the best of my knowledge and belief.
4. I/ we/ am are well aware of the fact that furnishing of any false information/ fabricated document would lead to rejection of my tender at any stage besides liabilities towards prosecution under appropriate law.
5. Each page of the tender document and papers submitted by my Company is authenticated, sealed and signed, and I take full responsibility for the entire documents submitted.

Date: -----

Place: -----
Authorized Person

Signature of the

Full Name: -----

Company Seal: -----

DECLARATION SHEET

We, _____ hereby certify that all the information and data furnished by our organization with regard to this tender specification are true and complete to the best of our knowledge. I have gone through the specification, conditions and stipulations in details and agree to comply with the requirements and intent of specification.

This is certified that our organization has been authorized (Copy attached) by the OEM to participate in Tender. We further certified that our organization meets all the conditions of eligibility criteria laid down in this tender document. Moreover, OEM has agreed to support on regular basis with technology / product updates and extend support for the warranty.

The prices quoted in the financial bids are subsidized due to academic discount given to RGIPT, Jais, Amethi.

| | |
|---|--|
| We, further specifically certify that our organization has not been Black-Listed/ De Listed or put to any Holiday by any Institutional Agency/ Govt. Department/ Public Sector Undertaking in the last three years. | |
| 1. Phone | |
| 2. Fax | |
| 3. E-mail | |
| 4. Contact Person Name | |
| 5. Mobile Number | |
| 6. TIN Number | |
| 7. PAN Number | |
| 8. (In case of on-line payment of Tender Fees) UTR No. (For Tender Fee) | |
| 9. (In case of on-line payment of EMD) UTR No. (For EMD) | |
| 10. Kindly provide bank details of the bidder in the following format: a) Name of the Bank | |
| b) Account Number | |
| c) Kindly attach scanned copy of one Cheque book page to enable us to return the EMD to unsuccessful bidder | |

(Signature of the Tenderer)

Name:

Seal of the Company

NON-BLACKLISTING DECLARATION

It has been certified that all information provided in tender form is true and correct to the best of my knowledge and belief. No forged / tampered document(s) are produced with tender form for gaining unlawful advantage. We understand that RGIPT, Amethi is authorized to make enquiry to establish the facts claimed and obtain confidential reports from clients.

In case it is established that any information provided by us is false / misleading or in the circumstances where it is found that we have made any wrong claims, we are liable for forfeiture of EMD and or any penal action and other damages including withdrawal of all work/purchase orders being executed by us. **Further RGIPT, Amethi is also authorized to blacklist our firm/company/agency and debar us in participating in any tender/bid in future.**

I / We assure the Institute that neither I / We nor any of my / our workers will do any act/s which is improper / illegal during the execution in case the tender is awarded to us.

Neither I/We nor anybody on my/our behalf will indulge in any corrupt activities/practices in my/our dealing with the Institute.

Our Firm/Company/Agency is not been blacklisted or banned by any Central Government Department/ State Government Department, PSU, University, Autonomous Institute or Any other Govt. Organization.

Date

Signature of the Tenderer

Place

Stamp

Note: This certificate should be executed on duly notarized `100/- Non Judicial Stamp Paper.

List of Govt. Organization/Department/Institutes

| List of Government Organization/Department/Institutes for whom the Bidder has undertaken such work during last three years (must be supported with work orders) | | |
|--|-------------------------------|--------------------|
| Name of the Organization | Name of Contact Person | Contact No. |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |

(Signature of the Tenderer)

Name:

Seal of the Company



RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY

CHECKLIST FOR SUBMISSION OF DOCUMENTS FOR FINANCIAL BID

For Exemptions:-

1. For Customs duty exemption in terms of Govt. Notification No. 51/96- Customs dated 23.07.1996 & 28/03 Customs dated 01.03.2003 and Central Excise duty Exemption in terms of govt. Notification No. 10/97- Central Excise dated 01.03.1997 & 28/13- Central Excise dated 01.10.2013.
Institute DSIR Notification No. TU/V/RG-CDE(1083)/2016 dated:03.02.2017 Valid up to 31.08.2021
2. For GST Exemption Notification No. 47/2017- Integrated tax (Rate).

**The Company can bid in Foreign Currency or Indian Currency
in the given format**

The bidder may add rows as per requirement to include the prices of all components/ parts, warranties, installation etc. whichever applicable.

FINANCIAL BID – Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System

A. Financial Bid in Indian Currency (For indigenous equipment):

- I. Name of the Bidder/Tenderer:
Name & Model No of offered goods:**

Tender No.:

| Sl. No. | Details | Total Amount (In INR) |
|---|---|------------------------------|
| 1 | Basic Cost of Inductively Coupled Plasma Optical Emission Spectrophotometer (ICPOES) System with accessories defined in Technical Specifications Statement of the Tender | |
| 2 | Add: Cost of Warranty (2 Year) | |
| 3 | Add: Packing & Forwarding Charges, If any | |
| 4 | Add: Installation & Commissioning Charges, if any | |
| 5 | Add: Transportation Charges up to RGIPT, Jais, Amethi, if any | |
| 6 | Add: GST (Taxes) | |
| Total Price (Sum of Sl. No 1 to 6) F.O.R RGIPT, Jais, Amethi Price | | |

Note: The Institute will provide DSIR & GST Exemption Certificate.

II. For Local Supplies -

| Sl. No. | Details | Total Amount (In INR) |
|---|---|------------------------------|
| 1 | Cost of Local Supplies defined in Technical Specifications Statement of the Tender Document (Please mention the items) | |
| 2 | Add: Packing & Forwarding Charges, If any | |
| 3 | Add: Installation & Commissioning Charges, if any | |
| 4 | Add: Transportation Charges up to RGIPT, Jais, Amethi, if any | |
| 5 | Add: Cost of Warranty | |
| 6 | Add: GST (Taxes) | |
| Total Price (Sum of Sl. No 1 to 5) F.O.R RGIPT, Jais, Amethi Price | | |

- I. Total Bid price in Indian Currency (Equipment Cost with accessories):.....In words:.....
- II. Total Bid price in Indian currency (Local Supplies):.....
- III. In words.....

Grand Total (I+II) =

In words:

- a) Indian Agents Name & Address
- b) The cost of optional items shall be indicated separately
- c) In case the price quoted in Foreign Currency (Other than INR), for the purpose of comparison the Exchange Rate prevailing on the date of opening of Financial Bid will be considered.

Date

Signature of the Tenderer

Place

Stamp

B. Financial Bid in Foreign Currency (for Imported equipment) :

**Price Schedule Form: Price schedule for goods being offered from abroad
In currency other than INR (Indian Rupees)
(Separate form to be used for each item offered)
(ON THE LETTER HEAD OF THE COMPANY)**

I. Name of the Bidder/Tenderer:

Name & Model No of offered goods:

Tender No.:

(Separate Prices should be defined for each instrument)

| Sl. No. | Details (Basic Cost of Equipment with accessories) | Price per unit in (-) |
|---|---|-----------------------|
| A. | Country of origin | |
| | Payment in Foreign Currency | |
| 1 | C.I.F Price up to Indian Airport/port in foreign currency | |
| 2 | Total Price at Indian Airport /port (C.I.F) | |
| B. | Reimbursement of Expenses (in Indian Rupees only) | |
| 3 | **Custom Duty against DSIR certificate (Please mentioned % also) | |
| 4 | Custom clearance and other charges, if any (with breakup) | |
| 5 | Inland Charges for Insurance & transportation to RGIPT Jais, Amethi | |
| 6 | Installation, commissioning and training Charges, if any | |
| 7 | Cost of Warranty (2 Year) | |
| Total Price F.O.R RGIPT, Jais, Amethi Price (A + B) (Sum of Sl. No 1 to 7) | | |

**** RGIPT, Jais, Amethi have Custom/Excise Exemption Certificate.**

For Local Supplies -

| Sl. No. | Details | Total Amount (In INR) |
|---|---|------------------------------|
| 1 | Cost of Local Supplies defined in Technical Specifications Statement of the Tender Document (Please mention the items) In INR only. | |
| 2 | Add: Packing & Forwarding Charges, If any | |
| 3 | Add: Installation & Commissioning Charges, if any | |
| 4 | Add: Transportation Charges up to RGIPT, Jais, Amethi, if any | |
| 5 | Add: Cost of Warranty | |
| 6 | Add: GST (Taxes) | |
| Total Price (Sum of Sl. No 1 to 6) F.O.R RGIPT, Jais, Amethi Price | | |

- I. Total Bid price in **Foreign Currency** (Equipment Cost with accessories):.....
In words:.....
- II. Total Bid price in Indian currency (Local Supplies):.....
In Words:.....

a) Indian Agents Name & Address

b) The cost of optional items shall be indicated separately

c) In case the price quoted in Foreign Currency (Other than INR), for the purpose of comparison the Exchange Rate prevailing on the date of opening of Financial Bid will be considered.

Date

Signature of the Tenderer

Place

Stamp