

RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY, JAIS, AMETHI

CORRIGENDUM-1

Date: 15.11.2019

Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/35 in rgipt.euniwizarde.com portal & Institute Website for the Procurement of Small and Wide Angle X-ray Scattering (SAXS/WAXS).

Corrections/Addendum to SAXS/WAXS Technical Specifications

<u>Before Pre-bid meeting</u>	<u>After Pre-bid meeting</u>
The sample translation should be higher than 50 mm in both directions	Sample translation should be 30-50 mm or higher in both direction
All detector must be placed in air for high reliability	All detector must be placed in air/vacuum for high reliability

N.B.: A marking system will be used to evaluate the bids. The bid that receives more than the threshold mark fixed by the purchase committee will be considered further for finance bid.

The last date to receive the bids is extended till 28th Nov 2019 3pm & the Opening of Techno- Commercial bid will be at 4pm.

Dr. U. Ojha



M.S. 

Chairman, Institute Equipment Purchase Committee