



राजीव गाँधी पेट्रोलियम प्रौद्योगिकी संस्थान
(संसद के अधिनियम के अधीन स्थापित राष्ट्रीय महत्व का एक संस्थान)
जायस, अमेठी, उत्तर प्रदेश, भारत-229304

RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY
(An Institution of National Importance Established under an Act of
Parliament)
Jais, Amethi - 229304, Uttar Pradesh, India

QUOTATION ENQUIRY

Ref. No.: RGIPT/JAIS/PEGE/UG/2023-24/02

Dated: 17.08.2023

Last Date and time for the quote: 28.08.2023 till 13:00 hours in the Office of the Department of Department of Petroleum Engineering & Geoengineering, 5th Floor, AB-1, RGIPT, Jais, UP-229304.

Dear Sir/Madam,

Please submit your lowest quotation for supplying the under mentioned item. Quotation must reach us before the date marked above and should contain the following information:

1. Clearly mention the date of validity of offer.
- 2. Kindly clearly mention your E-mail ID and Mobile number.**
3. Kindly mention appropriate GST % as applicable for supplying of Goods & Services.
4. Please mention your GST registration number and PAN in the quotation.
5. Conditions of supply and terms of payment shall be clearly mentioned in the quotation.

Quotation must be sent in **sealed envelope** with word "**QUOTATION for "Slake Durability Tester"**" and addressed to Office of the Department of Petroleum Engineering & Geoengineering, 5th Floor, AB-1, RGIPT, Jais- 229304 (UP). Our reference number and last date as given above should be clearly marked over it.

Details of the item are as follows:

Item Name: **Slake Durability Tester**

Quantity: 01 Nos.

Technical Specification of "Slake Durability Tester"

Technical Specifications		Bidder's Specification	Compliance/ Deviation
Revolutions per Minute	20 RPM		
Mesh Size	2 mm		
Drum requirement	4 Drums		
Drum Dimension (diameter)	140 mm		

Drum Dimension (height)	100mm		
A digital timer automatically stops the motor after the preset time	Yes		
Product Dimensions	48 x 14 x 9.25in (1,219 x 355 x 235mm), WxDxH		
Followed Standard	ASTM D4644		
Net weight	30 Kg or more		
Thermal protection	Yes		
Warranty & Maintenance	<p>Minimum 3 year's OEM warranty after successful delivery & installation</p> <ul style="list-style-type: none"> • Manufacturer/authorized suppliers should have at least ISO certificate • A soft copy and hard copy of the operation and maintenance manuals in English must be provided. 		

Note:

- The machine mounts a digital display for the visualization of the residual testing time (10 minutes for ASTM D4644 standard).
- The water level is indicated by a mark on the container.
- All charges (packing, freight) including 18% GST should be mentioned in detail.
- The company must provide **a Price reasonability certificate** on letterhead



Prof. Satish Kumar Sinha
Chairman, DPC
Department of Petroleum Engineering & Geoengineering,
RGIPT, Jais