



राजीव गाँधी पेट्रोलियम प्रौद्योगिकी संस्थान
(संसद के अधिनियम के अधीन स्थापित राष्ट्रीय महत्व का एक संस्थान)

जायस ,अमेठी, उत्तर प्रदेश, भारत-229304

RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY
(An Institution of National Importance Established under an Act of Parliament)

Jais, Amethi - 229304, Uttar Pradesh, India

QUOTATION ENQUIRY

Ref. No.: RGIPT/JAIS/PEGE/UG/QE/2023-24/09

Dated: 24.11.2023

Last Date and time for the quote: 04.12.2023 till 18:00 hours in the Office of the Department of Department of Petroleum Engineering & Geoengineering, 5th Floor, AB-1, RGIPT, Jais, UP-229304.

Dear Sir/Madam,

Please submit your lowest quotation for supplying the under mentioned item. Quotation must reach us before the date marked above and should contain the following information:

1. Clearly mention the date of validity of offer.
2. **Kindly clearly mention your E-mail ID and Mobile number.**
3. Kindly mention appropriate GST % as applicable for supplying of Goods & Services.
4. Please mention your GST registration number and PAN in the quotation.
5. Conditions of supply and terms of payment shall be clearly mentioned in the quotation.

Quotation must be sent in **sealed envelope** with word **“QUOTATION for “Heating jacket for liquid Permeameter”** and addressed to Office of the Department of Petroleum Engineering & Geoengineering, 5th Floor, AB-1, RGIPT, Jais- 229304 (UP). Our reference number and last date as given above should be clearly marked over it.

Details of the item are as follows:

Item Name: **Heating jacket for liquid Permeameter**

Quantity: **01 No.**


Technical Specifications - “Heating jacket for liquid permeameter”

Specifications	Bidder's Specification	Compliance/ Deviation
Heating jacket for liquid Permeameter • Two Nos. for Accumulator. Ceramic Band Heater 250W, 240VAC SS braiding protection Heater Dimensions- ID 65, Width 150 Maximum Power Rating 300W Power supply 240 VAC High temperature fibre glass lead wire • Two Nos. for Core Holder Ceramic Band Heater 250W, 240VAC SS braiding protection Heater Dimensions- ID 95, Width 90 Maximum Power Rating 250W		

Power supply 240 VAC		
High temperature fibre glass lead wire		
Warranty Minimum one Year		

Note:

- All charges (packing, freight, installation, and commissioning) including GST should be mentioned in detail.
- The company must provide a Price reasonability certificate on letterhead.
- Company must provide a copy of PAN and GST Registration, and Cancelled Cheque.



Prof. Satish Kumar Sinha
Chairman, DPC
Department of Petroleum Engineering & Geoengineering,
RGIPT, Jais