



राजीव गाँधी पेट्रोलियम प्रौद्योगिकी संस्थान
(संसद के अधिनियम के अधीन स्थापित राष्ट्रीय महत्व का एक संस्थान)
जायस, अमेठी, उत्तर प्रदेश, भारत-229304

RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY
(An Institution of National Importance Established under an Act of Parliament)
Jais, Amethi - 229304, Uttar Pradesh, India

QUOTATION ENQUIRY

Ref. No.: RGIPT/JAIS/PEGE/UG/QE/2023-24/13

Dated: 2.02.2024

Last Date and time for the quote: 12.02.2024 till 17:00 hours in the Office of the Department of Department of Petroleum Engineering & Geoengineering, 5th Floor, AB-1, RGIPT, Jais, UP-229304.

Dear Sir/Madam,

Please submit your lowest quotation for supplying the under mentioned item. Quotation must reach us before the date marked above and should contain the following information:

1. Clearly mention the date of validity of offer.
2. **Kindly clearly mention your E-mail ID and Mobile number.**
3. Kindly mention appropriate GST % as applicable for supplying of Goods & Services.
4. Please mention your GST registration number and PAN in the quotation.
5. Conditions of supply and terms of payment shall be clearly mentioned in the quotation.

Quotation must be sent in **sealed envelope** with word "**QUOTATION for "Multiparameter Analyzer for Reservoir formation fluids"**" and addressed to Office of the Department of Petroleum Engineering & Geoengineering, 5th Floor, AB-1, RGIPT, Jais- 229304 (UP). Our reference number and last date as given above should be clearly marked over it.

Details of the item are as follows:

Item Name: **Multiparameter Analyzer for Reservoir formation fluids**

Quantity: **01 No.**

Technical Specifications - "Multiparameter Analyzer for Reservoir formation fluids"


| Specifications | Bidder's Specification | Compliance/ Deviation |
|--|------------------------|-----------------------|
| <ul style="list-style-type: none">• Structural fabrications should meet the following requirements:• Technical Specification:• Table-top/Bench-top• Safe & comfortable• Compact and Lightweight• Calibration certificates should be provided accordingly• LCD display with different modes button• Limit alarm alert when out of range• Auto logging function for continuous monitoring• Reset function should be provided • pH• Range: - -2.00 to 20.00 pH• Resolution: - 0.01 PH or better• Accuracy: - 0.002pH • ORP• Range: - +/- 2000.00mV• Resolution: - 0.1mV | | |

[Handwritten signature]

| | | |
|--|--|--|
| <ul style="list-style-type: none"> • ION • Concentration: - 0.001 to 20000ppm • Resolution: - 2 or 3 digit • Accuracy: - 0.5% full scale (Monovalent), 1% full scale (divalent) • Conductivity • Range: - 0.05μS to 500mS • Resolution: - 0.01/0.1μS; 0.001/0.01/0.1mS • Accuracy: - +/- 1% full scale • TDS • Range: 0.05ppm to 500ppt (@ TDS factor 1.00) • Resolution: - 0.01/0.1 ppm; 0.001/0.01/0.1 ppt • Accuracy: - +/- 1% full scale. • Salinity • Range: 0 to 80ppt • Resolution: - 0.01/0.1 ppm; 0.001/0.01/0.1 ppt • Accuracy: - +/- 1% full scale • Temperature • Range: 0 to 100°C • Resolution: - 0.1°C • Accuracy: - +/- 0.1 to 0.3 °C • Resistivity • Range: 2.00Ω to 20.00MΩ • Resolution: - 0.01/0.1 Ω; 0.001/0.01 kΩ; 0.01 MΩ • Accuracy: - +/- 1% full scale | | |
|--|--|--|

Note:

- All charges (packing, freight, installation, and commissioning) including GST should be mentioned in detail.
- Company must provide a copy of PAN and GST Registration, and Cancelled Cheque.
- Bidder must provide product brochure.


21/2/24

Prof. Satish Kumar Sinha
Chairman, DPC
Department of Petroleum Engineering & Geoengineering,
RGIPT, Jais