



RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY
(An Institution of National Importance Established under an Act of Parliament)
Jais, Amethi - 229304, Uttar Pradesh, India

QUOTATION ENQUIRY

Ref. No.: RGIPT/JAIS/S&H/MMA/2024-25

Dated: 18.10.2024

Last Date and time for the quote: 1.11.2024 till 17:00 hours in the Office of the Department of Science & Humanities, 5th Floor, AB-2, RGIPT, Jais, UP-229304.

Dear Sir/Madam,

Please submit your technical and financial bid quotation in two separate envelopes for supplying the under-mentioned item. Quotation must reach us before the date marked above and should contain the following information:

1. Clearly mention the date of validity of offer.
2. **Kindly clearly mention your E-mail ID and Mobile number.**
3. Kindly mention appropriate GST % as applicable for supplying of Goods & Services.
4. Please mention your GST registration number and PAN in the quotation.
5. Conditions of supply and terms of payment shall be clearly mentioned in the quotation.
6. Please provide the technical specifications and the bill in two separate envelopes.
7. **In case technical and financial quotes are not received in two separate envelopes. The bid would be deemed to be cancelled/Invalid.**

Quotation must be sent in **sealed envelope** with word **“QUOTATION for “Mass Measurement Apparatus”** and addressed to Office of the **Department of Sciences & Humanities, 5th Floor, AB-2, RGIPT, Jais, UP-229304**. Our reference number and last date as given above should be clearly marked over it.

Details of the item are as follows:

Item Name: Mass Measurement Apparatus

Quantity: 01 No.

Technical Specifications – “Mass Measurement Apparatus”

Specifications	Bidder's Specification	Compliance/ Deviation
Mass Measurement Apparatus Max. (Capacity): 220g • Readability (d): 0.0001g • Repeatability (+): 0.2mg • Min. Capacity: 10mg • Calibration: Internal Calibration • Display: 5 inch Touch Screen • Response Time: 3 to 5 Sec • Pan Size: 80mm Φ • Weighing Chamber: 240 x 190 x 265 mm • Unit: mg, g, ct, oz • Power Supply: AC 220V / 50-60 Hz • Dimension (LxWxH): 475 x 315 x 485 mm • Tare Range: Full • Sensitivity Drift: + 2ppm / C		



RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY
(An Institution of National Importance Established under an Act of Parliament)
Jais, Amethi - 229304, Uttar Pradesh, India

Warranty: 24 months Delivery: Within 20 – 45 days after receiving the confirmed order		
--	--	--

Note:

- All charges (Packing, Freight, Installation, and Commissioning) including GST should be mentioned in detail.
- Company must provide a copy of PAN and GST Registration, and Cancelled Cheque.
- Bidder must provide product brochure.

Atul Sharma
18/03/24

Prof. Atul Sharma
Chairman, HOD S&H
Department of Sciences & Humanities,
RGIPT, Jais-229304